
FICHA TECNICA

Nombre del producto	Medidor de presión de aceite relleno de glicerina de cáscara de acero inoxidable 304
Caso de material	304 de acero inoxidable
Interior	De aleación de latón
Servicio destinado a	Adecuado para usar con agua, aceite, aire, gas u otra versión de llenado de líquidos no corrosivos para condiciones de servicio adversas donde existe pulsación o vibración.
Escala estándar	0-100bar 0-1500psi
Ventana	PC
La clase de precisión	1.60%
Conexión	1/4 "NPT
Tamaño Nominal	2,5 "(Ø = 63mm)
De medio	Glicerina











CERTIFICACION



INGEER CERTIFICATION ASSESSMENT SERVICES

MANAGEMENT SYSTEM CERTIFICATE

Certificate No.: 117 19 QU 0053-05 R0M

This is to certify the quality management systems of

GARHOPE MEASUREMENT CO., LTD

Unified Social Credit Code 91321000778005187K

Location (No. 61, South Huanghe Road), Private Industrial Park II Phase, New Area, Jiangdu District, Yangzhou, Jiangsu

has been assessed and registered as meeting the requirements of

GB/T19001-2016/ISO9001:2015

Scope of approval

Production of 2.5 Grade Common Pressure Gauge (Y40, Y50, Y60), 1.6 Grade Common Pressure Gauge (Y100); 2.5 Grade Vibration-proof Pressure Gauge (Y40), 1.5 Grade Vibration-proof Pressure Gauge (Y50, Y60, Y100), WSS-100 (1.5 Grade) Pressure Gauge, Bimetal Thermometer (Excluding the Products Refer to CMC License, Only for Export)

Signed by:

First Certification: 07 May 2019
Expiry Date: 06 May 2022

Shanghai Ingeer Certification Assessment Co.,Ltd.
Certification and Accreditation Administration of PRC:CNCA-R-2003-117
Tel: 400-182-9001/+86 21-51114700
Web: www.icas.org.cn
Add: Room 801,HuaDing Mansion, 2368# West Zhongshan Rd., Xuhui District, Shanghai, China, 200235

Focus on icas WeChat platform

First Surveillance Audit | Second Surveillance Audit | Third Surveillance Audit

The ownership of the certificate belongs to Shanghai Ingeer Certification Assessment Co., Ltd. The information & validation of this certificate can be checked on the CNCA website: WWW.CNCA.GOV.CN and ICAS website, or by calling ICAS's clients services Dept. The certificate is only valid when used together with related permits when appropriate. If the organization can't effectively maintain the above management system, ICAS has the right to withdraw the qualification certificate.